

<b>Topic</b>	Jamaican approval of Colilert <sup>®</sup> , Colilert <sup>®</sup> -18, Colisure <sup>®</sup> Quanti-Tray <sup>®</sup> /Quanti-Tray <sup>®</sup> /2000 for the testing of all waters
<b>Source</b>	Director Environmental Health Unit, Ministry of Health & Environment; Kingston, Jamaica
<b>Date</b>	December 5, 2008

## **Highlights:**

- Colilert, Colilert-18, Colisure, Quanti-Tray and Quanti-Tray/2000 are approved for the detection of Total coliforms and *E coli* in all waters in Jamaica
- Acceptance is based on inclusion in Standard Methods for the Examination of Water and Wastewater, current edition
- Approval recognition was received by direct, personal communication from the Director of the Environmental Health Unit of the Ministry of Health & Environment in Kingston, Jamaica
- Laboratories must have a color comparator in-house when using either Colilert or Colilert-18