

Topic	Jamaican approval of Colilert [®] , Colilert [®] -18, Colisure [®] Quanti-Tray [®] /Quanti-Tray [®] /2000 for the testing of all waters
Source	Director Environmental Health Unit, Ministry of Health & Environment; Kingston, Jamaica
Date	December 5, 2008

Highlights:

- Colilert, Colilert-18, Colisure, Quanti-Tray and Quanti-Tray/2000 are approved for the detection of Total coliforms and *E coli* in all waters in Jamaica
- Acceptance is based on inclusion in Standard Methods for the Examination of Water and Wastewater, current edition
- Approval recognition was received by direct, personal communication from the Director of the Environmental Health Unit of the Ministry of Health & Environment in Kingston, Jamaica
- Laboratories must have a color comparator in-house when using either Colilert or Colilert-18